

Performance of Two Rapid Antigen Detection Tests for Detecting COVID-19 Compared to RT-PCR in Indonesia

by Universitas Warmadewa Admin

Submission date: 13-Dec-2023 09:13AM (UTC+0700)

Submission ID: 2257397177

File name: BPJ_Vol_15_No_3_p_1271-1275.pdf (247.08K)

Word count: 2620

Character count: 14383

Performance of Two Rapid Antigen Detection Tests for Detecting COVID-19 Compared to RT-PCR in Indonesia

Anak Agung Gede Budhitesna^{1,2*}, Dewa Putu Surawan^{1,3},
Ratna Kartikadewi¹, Adi Harta Yoga¹, Putri Permana Lestari¹,
Nyoman Trisna Sumadewi¹ and Sri Masyeni^{1,3}

¹Department of Internal Medicine, Faculty of Medicine and Health Sciences, Universitas Warmadewa, Denpasar, Bali, 80235, Indonesia.

²Department of Internal Medicine, Sanjiwani Hospital, Gianyar, Bali, 80235, Indonesia.

³Department of Internal Medicine, Tabanan General Hospital, Tabanan, Bali, 82121, Indonesia.

*Corresponding Author E-mail: agbudhitesna@yahoo.com

<https://dx.doi.org/10.13005/bpj/2463>

(Received: 29 April 2022; accepted: 29 August 2022)

Antigen tests to screen coronavirus disease 2019 (COVID-19) is effective in symptomatic patients, leading to its wide usage in informing whether the person is COVID-19 positive or negative. Our current work had an objective to investigate the diagnostic performance of two antigen-detecting rapid diagnostic tests (Ag-RDTs) which are commonly used in Indonesia. A cross-sectional study was carried out to compare specificity, sensitivity, as well as expected predictive values of Anhui Ag-RDT and Lungene Ag-RDT by comparing the results with that obtained from real-time reverse transcription-polymerase chain reaction (RT-PCR) assay. A total of 98 samples were tested for both Ag-RDTs and RT-PCR. The median value of the patients age obtained to be 41.78 years old (interquartile range: 1 to 91 years old). The proportion between female and males was: 52.53% vs 47.47%. The sensitivities of Anhui Ag-RDT and Lungene Ag-RDT were 55.56% and 51.58%, where both Ag-RDTs had specificity of 100%. In conclusion, sensitivity values of Lungene Ag-RDT and Anhui Ag-RDT are similar, where both possess 100% specificity with zero false-positive results. Both of the investigated Ag-RDTs are useful since positive results are likely to be COVID-19 positive.

Keywords: COVID-19; Ct value; Diagnostic Performance; Rapid Antigen Test; RT-PCR.

As of November 2021, coronavirus disease 2019 (COVID-19) has impacted hundreds of millions of people around the world, and is responsible for more than 5.2 million deaths since its first discovery in December 2019¹. Many countries in the world have been reported the second wave of this severe acute respiratory cases and imposed the national lockdown policy. Of which, Indonesia is also encountering the threat of COVID-19 cases. More than 4.2 millions

of Indonesian were infected by severe acute respiratory syndrome corona virus 2 (SARS-CoV-2), resulting more than 143,000 deaths as of November 2021². Many efforts have been employed to reduce the number of COVID-19, such as enforcing health protocol law and distributing the COVID-19 vaccine nationally². Unfortunately, there is still no specific anti SARS-CoV-2 drugs have been developed and mutated SARS-CoV-2 keeps spreading. In this case, prompt diagnosis

and identification of COVID-19 patients who are vulnerable for developing severe courses are important.

The officially recognized diagnosing method for COVID-19 is nucleic acid analysis with real-time reverse transcription-polymerase chain reaction (RT-PCR) as a direct detection method. Although RT-PCR owes a high diagnostic performance for both symptomatic and asymptomatic cases, based on a public health perspective, RT-PCR has some drawbacks such as no optimal turnaround time³, high operational cost, and requiring skilled operator. To answer these challenges, antigen-detecting rapid diagnostic test (Ag-RDT) is employed, where it does not require laboratory capacities and skilled personnel. Moreover, Ag-RDT has a logistic advantage since the diagnostic test could be performed *in-situ*. However, there is a need to investigate the diagnostic performance of Ag-RDT. Hence, in this study, we have compared the results obtained from Ag-RDT with that of the gold standard test (RT-PCR) to determine its sensitivity and specificity.

METHODS

Study designs and ethical approval

A cross-sectional prospective study was conducted at the regional reference COVID-19 diagnostic laboratory, located at Universitas of Warmadewa, Denpasar – Indonesia. Specimens for the Ag-RDT and RT-PCR analysis were acquired by means of nasopharyngeal swab, between June and July 2021. Flocked probe was utilized for the nasopharyngeal swab to collect the specimen, and subsequently dissolved in the universal transport medium (UTM, Copan Diagnostics Inc, US). Specimens were tested before 2 hours had elapsed from their receipt in the laboratory. The ethical clearance had been granted by the Faculty of Medicine and Health Sciences, Universitas of Warmadewa (052/Unwar/FKIK/EC-KEPK/VI/2021).

Real-time reverse transcription quantitative polymerase chain reaction

To screen the presence of SARS-CoV-2, the specimen was analyzed using the extraction-free RT-PCR on AccuPrep® Viral RNA kit extraction (Bioneer Inc., Oakland), and nBioCoV-19 RT-PCR kit (Biofarma Inc.,

Indonesia), where the protocol follows that of the manufacturer's. The specimens were diagnosed for the viral presence through a single-step multiplex RT-PCR on Bio-Rad CFX96™ thermal cycler (CA, US). Amplifications of the viral gene were then observed based on FHEX (internal control), Cal Red 610 (RNA-dependent RNA-polymerase gene), and AM (E gene) fluorophores. Samples showing cycle threshold (Ct) values < 40 were assigned as COVID-19 positive.

Antigen-detecting rapid diagnostic tests

Two commercial Ag-RDTs included in this study: Anhui (Anhui Deep Blue Medical Technology Co., Ltd, China), and Lungene (Hangzhou Clongene Biotech Co, Ltd, China). Diagnosis using Ag-RDTs followed the instruction provided by the manufacturer.

Data analysis

Results from the Ag-RDTs were compared with that of RT-PCR, which was assigned as a reference standard by the Foundation for Innovative New Diagnostics⁴. Sensitivity along with specificity of both investigated Ag-RDTs were taken as a parameter for their diagnostic performance. Youden's J statistic was employed to calculate the cut-off of the Ct value with highest sensitivity and specificity. By employing the variation of hypothetical positivity ranged between 0.5 and 20%, expected negative (NPV) along with positive (PPV) predictive values were quantified, following suggestions from previous reports⁴⁻⁶.

RESULTS

Demographic data of the subjects (n=98) along with their test results using RT-PCR and two commercially Ag-RDTs (Anhui and Lungene) have been presented (Table 1). This study was participated by male (47.47%) and female (52.53%) subjects with median age of 41.78 years old. As many as 27.27% of the samples were positive by RT-PCR, 84.85 – by Lungene Ag-RDT, and 85.86% – by Anhui Ag-RDT. The mean Ct value of the positive sample was 33.4.

Sensitivity and specificity of Lungene Ag-RDT and Anhui Ag-RDT have been presented in Table 2. The percentage of false-negative results reached 12% in Lungene Ag-RDT, whilst false-positive results were not found. The collective specificity as well as sensitivity of Lungene Ag-

RDT were therefore 55.6% and 100%, respectively. Meanwhile, 51.8% sensitivity and 100% specificity were obtained by Anhui Ag-RDT. The accuracy percentages for Lungene and Anhui Ag-RDTs were 87.9 and 86.9, respectively. The expected NPVs were obtained to be 99.9%, 99.8%, 98.9%, 97.7% and 94.9% with a variation of positivity rates (0.5%, 1%, 5%, 10% and 20%, respectively). Meanwhile, PPV remained constant 100% at all positivity rates. The optimum value of the Ct cut-off for both investigated Ag-RDTs was 29.

DISCUSSION

The present study compared the sensitivity and specificity between two RDT-Ag tests to screen the availability of the COVID-19 causing virus on the suspected specimen. Lungene Ag-RDT had a higher sensitivity than Anhui Ag-RDT, with 55.6% and 51.8%, respectively. However, the specificity of the two assays was identical (100%) with zero false-positive. These assays are sub-optimally because the ideal test should have a sensitivity of >95% and a specificity of 100%⁷. WHO has set a minimal sensitivity requirement of 80% and a specificity of 97% as acceptable tests⁸. A comprehensive review has documented variations in the estimated sensitivity as well as specificity of the diagnostic accuracy of Ag-RDT assays compared to RT-qPCR⁵. The sensitivity and specificity vary depending on the brand in which

the sensitivity and the and specificity of the tests in asymptomatic individuals ranged between 29-85% and 14-100%, respectively⁵.

Previous studies found that the diagnostic tests using Ag-RDT have high sensitivity in those with higher viral loads^{5, 9}. For example, the Ag-RDT sensitivity compared to NAAT was higher in symptomatic patients than in those who are asymptomatic (66.7% vs. 47.6%)⁹. A review found that the average sensitivity of the 48 included studies was 72.0% (95%CI: 63.7%, 79.0%). The mean sensitivity decreased over time following the symptoms onset and was lower in the second week (51.0%; 95%CI: 40.8%, 61.0%) than in the first (78.3%; 95%CI: 71.0%, 84.1%). Better sensitivity was obtained in individuals with a higher load of the virus (Ct value < 25) compared to those with a lower viral load (94.5% vs 40.7%)⁵.

In our study, the sensitivity of Lungene Ag-RDT was 55.6%, in which 12 of the 84 individuals who were recorded as negative by the assay turned out to be positive after tested by RT-PCR. Similarly, 13 of the 85 negative tested by Anhui Ag-RDT samples were positive after tested by RT-PCR, indicating that 12.3% of the samples were false-negative. The sensitivity of these Ag-RDT tests might be influenced by the Ct value of the samples. According to Bruzzonea, the question of which RT-PCR cut-off should be employed for deciding whether a sample infected with SARS-CoV-2 or not is still being debated¹⁰. A study suggested that high viral load was obtained when the Ct value was < 24, while Ct values above 24 indicated a significantly reduced infectiousness¹¹. Meanwhile, according to another study summarized in Jefferson's review, a more conservative limit for acceptable RT-qPCR results is in the Ct value range

Table 1. Demographic data of the subject (n=98)

| Variable | Number | Frequency (%) |
|-----------------|---------------|---------------|
| Age | | |
| Min – Max | 1 - 91 | |
| Median (SD) | 41.78 (19.12) | |
| Gender | | |
| Male | 47 | 47.47 |
| Female | 52 | 52.53 |
| Antigen Lungene | | |
| Negative | 84 | 84.85 |
| Positive | 15 | 15.15 |
| Antigen Anhui | | |
| Negative | 85 | 85.86 |
| Positive | 14 | 14.14 |
| RT-PCR | | |
| Negative | 72 | 72.73 |
| Positive | 27 | 27.27 |

Table 2. Diagnostic performance of Ag-RDTs as compared with RT-PCR (n=98)

| Diagnostic test | Lungene Ag-RDT | Anhui Ag-RDT |
|--------------------|----------------|--------------|
| Sensitivity (%) | 55.6 | 51.8 |
| Specificity (%) | 100 | 100 |
| True negative (%) | 72 | 72 |
| False negative (%) | 12 | 13 |
| True positive (%) | 15 | 14 |
| False positive (%) | 0 | 0 |
| Accuracy (%) | 87.9 | 86.9 |

of 24 to 35¹². In our present research, the optimal Ct value was observed at the Ct cut-off of 29 with mean Ct value of 33.4 obtained from the positive samples. There are some reasons for the false negative in our samples: (a) some samples might have high Ct values making the viral load was low; or (b) the antigen of the viruses in the samples might had low affinity with antibody used in the tested kit due to mutations of the viruses. To be an accurate testing, the false negative should be low to ensure to be able to detect or diagnose as many as the disease in the population^{13,14}.

Our results suggested that despite low sensitivity of Lungene Ag-RDT and Anhui Ag-RDT, these assays still can be used even they do not meet WHO criteria. This is because a positive rapid test sample is most likely to be identified as true positive without further confirmation by RT-PCR and this could help to ease the pressure on health facilities, in particular those with limited resources. As for the negative test, we recommend that RT-PCR need to be tested if the patients showing presumptive for COVID-19. This is in line with the European Center for Disease Prevention and Control's recommendations⁶. Rapid antigen testing is essential in everyday life because it is simple to use, does not necessitate special skills, and has high specificity.

CONCLUSION

Lungene Ag-RDT and Anhui Ag-RDT have similar sensitivity, 55.6% and 51.8%, respectively; however, there was no difference in specificity (100%) with zero false-positive results. Although less than optimal, antigen diagnostics using Lungene Ag-RDT and Anhui Ag-RDT are still important since positive results are mostly positive for COVID-19 and this could help for rapid management of the patients in particular in point-of-care facilities with limited resources.

Conflict of interest

The authors declare no conflict of interest.

Funding

This study received no additional funding.

Author contributions

All authors contributed equally to the study, including in data collection, statistical analysis, and data synthesis.

REFERENCES

1. World Health Organization. WHO coronavirus (COVID-19) dashboard 2021; Available from: <https://covid19.who.int> (Accessed on: 1 February 2022).
2. World Health Organization. Update on coronavirus disease 2021; Available from: <https://www.who.int/indonesia/news/novel-coronavirus> (Accessed on: 1 February 2022).
3. Russo A, Minichini C, Starace M, Astorri R, Calo F, Coppola N, et al. Current Status of Laboratory Diagnosis for COVID-19: A Narrative Review. *Infect Drug Resist.*; **13**:2657-65 (2020). PubMed PMID: 32801804. Pubmed Central PMCID: PMC7413717. Epub 20200803.
4. Foundation for Innovative New Diagnostics (FIND). Comparative evaluation of lateral flow assay tests that directly detect antigens of SARS-CoV-2020 [cited 2021 29 March]. Available from: <https://www.finddx.org/wp-content/uploads/2020/04/20200421-COVID-Ag-RDT-Evaluation-Synopsis.pdf>. (Accessed on: 1 February 2022).
5. Dinnes J, Deeks JJ, Berhane S, Taylor M, Adriano A, Davenport C, et al. Rapid, point-of-care antigen and molecular-based tests for diagnosis of SARS-CoV-2 infection. *Cochrane Database Syst Rev.*; **3**:CD013705 (2021). PubMed PMID: 33760236. Pubmed Central PMCID: PMC8078597. Epub 20210324.
6. European Centre for Disease Prevention and Control. Options for the use of rapid antigen tests for COVID-19 in the EU/EEA and the UK2020 [cited 2021 29 March]. Available from: <https://www.ecdc.europa.eu/sites/default/files/documents/options-use-of-rapid-antigen-tests-for-COVID-19.pdf>. (Accessed on: 1 February 2022).
7. Nalumansi A, Lutalo T, Kayiwa J, Watera C, Balinandi S, Kiconco J, et al. Field evaluation of the performance of a SARS-CoV-2 antigen rapid diagnostic test in Uganda using nasopharyngeal samples. *International Journal of Infectious Diseases.*; **104**:282-6 (2021).
8. World Health Organization. Antigen-detection in the diagnosis of SARS-CoV-2 infection. 2021; Available from: <https://www.who.int/publications/i/item/antigen-detection-in-the-diagnosis-of-sars-cov-2-infection-using-rapid-immunoassays>. (Accessed on: 1 February 2022).
9. Kahn M, Schuierer L, Bartenschlager C, Zellmer S, Frey R, Freitag M, et al. Performance of antigen testing for diagnosis of COVID-19: a direct comparison of a lateral flow device to nucleic acid amplification based tests. *BMC*

10. Bruzzone B, De Pace V, Caligiuri P, Ricucci V, Guarona G, Pennati BM, et al. Comparative diagnostic performance of rapid antigen detection tests for COVID-19 in a hospital setting. *International Journal of Infectious Diseases.*; **107**:215-8 (2021).
11. Bullard J, Dust K, Funk D, Strong JE, Alexander D, Garnett L, et al. Predicting Infectious Severe Acute Respiratory Syndrome Coronavirus 2 From Diagnostic Samples. *Clinical Infectious Diseases* ;**71**(10):2663-6 (2020).
12. Jefferson T, Spencer EA, Brasse J, Heneghan C. Viral Cultures for Coronavirus Disease 2019 Infectivity Assessment: A Systematic Review. *Clinical Infectious Diseases.*; **73**(11):e3884-e99 (2021).
13. Palupi IR, Sugianto P. Accuracy of Meningitis Bacterial Score (MBS) as an indicator in establishing the diagnosis of acute bacterial meningitis in adults. *Narra J.*; **1**(3) (2021).
14. Chaudhary R, Bhatta S, Singh A, Pradhan M, Shrivastava B, Singh YI, et al. Diagnostic performance of GeneXpert MTB/RIF assay compared to conventional Mycobacterium tuberculosis culture for diagnosis of pulmonary and extrapulmonary tuberculosis, Nepal. *Narra J.*; **1**(2) (2021).

Performance of Two Rapid Antigen Detection Tests for Detecting COVID-19 Compared to RT-PCR in Indonesia

ORIGINALITY REPORT

4%

SIMILARITY INDEX

4%

INTERNET SOURCES

1%

PUBLICATIONS

3%

STUDENT PAPERS

MATCH ALL SOURCES (ONLY SELECTED SOURCE PRINTED)

2%

★ doaj.org

Internet Source

Exclude quotes On

Exclude matches < 1%

Exclude bibliography On